



SHIMADZU

LabSolutions

# Analysis Report

## <Sample Information>

Sample Name : Rice\_2

Sample ID : Rice\_5.0ng-g

Data Filename: Rice Accuracy\_2.lcd

Method Filename : Bahaudeen calibration method.lcm

Batch Filename : Accuracy\_02072020-1.lcb

Vial # : 1-16

Sample Type : Spiked


Injection Volume : 10 uL

Date Acquired : 2/7/2020 7:37:50 PM Acquired by : Analytical Laboratory Date Processed :

21/8/2020 9:36:33 AM Processed by : Analytical Laboratory

## <Chromatogram>

mV



## <Peak Table>

Detector A Ex:360nm,Em:440nm

Peak#	Ret. Time	Conc.	Name
1	11.045	4.914	AFG2
2	11.878	0.000	
3	12.527	4.946	AFG1
4	14.649	4.976	AFB2
5	15.669	0.000	
6	16.959	4.948	AFB1
7	17.735	0.000	
8	18.850	0.000	